# MEASURING THE NIGHT SKY BRIGHTNESS WITH THE LIGHTMETER

A. Müller, 1,2 G. Wuchterl, 3 and M. Sarazin 1

### RESUMEN

Favor de proporcionar un resumen en español. If you are unable to translate your abstract into Spanish, the editors will do it for you. We present a newly developed, low-cost photometer for long-term monitoring of the night sky brightness and light pollution on Earth. The so-called LIGHTMETER is an as far as possible stand-alone operational, fully weatherproof, and maintenance-free device. It provides a high data sampling rate of up to 1 Hz as well as a superb sensitivity covering the whole brightness range down to the darkest night time conditions. The excellent performance of the LIGHTMETER allows a continuously monitoring of the night sky brightness and opens a wide range of applications at an observatory site like determining overall sky conditions in real time, cloud detection and estimation of their velocity, measuring relative changes in extinction as well as the detection of long term trends in brightness caused by an increase of artificial illumination. We will present first results of measurements taken at Cerro Armazones, one of the best observing sites in the world and the selected site of the planned European Extremely Large Telescope (E-ELT).

#### ABSTRACT

We present a newly developed, low-cost photometer for long-term monitoring of the night sky brightness and light pollution on Earth. The so-called LIGHTMETER is an as far as possible stand-alone operational, fully weatherproof, and maintenance-free device. It provides a high data sampling rate of up to 1 Hz as well as a superb sensitivity covering the whole brightness range down to the darkest night time conditions. The excellent performance of the LIGHTMETER allows a continuously monitoring of the night sky brightness and opens a wide range of applications at an observatory site like determining overall sky conditions in real time, cloud detection and estimation of their velocity, measuring relative changes in extinction as well as the detection of long term trends in brightness caused by an increase of artificial illumination. We will present first results of measurements taken at Cerro Armazones, one of the best observing sites in the world and the selected site of the planned European Extremely Large Telescope (E-ELT).

Key Words: instrumentation: photometers — site testing — light pollution — methods: observational — methods: data analysis

#### 1. INTRODUCTION

Monitoring the brightness of the night sky is usually an elaborate and expensive task. The operation of complex all-sky cameras and telescope systems make continuous measurements at any time and weather condition impossible. In addition, these systems come along with a complex infrastructure like buildings in order to protect the sensitive hardware against weather and solar illumination. We present a newly developed photometer, which we call LIGHTMETER in the following. The LIGHTMETER consists of a photo voltaic cell, electronic components like amplifier and A/D converter, which are imple-



<sup>&</sup>lt;sup>2</sup>Max Planck Institute for Astronomy, Königstuhl 17, D-69117 Heidelberg, Germany (amueller@mpia.de).



Fig. 1. The LIGHTMETER device.

mented in an aluminum frame and can be connected to any computer using USB (Figure 1). In order to be able to integrate over the full hemisphere of the sky, the Lightmeter has to be assembled parallel to the horizon, preferable above any structure to avoid dome shadowing and direct illumination by artificial

<sup>&</sup>lt;sup>3</sup>Thüringer Landessternwarte Tautenburg, Sternwarte 5, D-07778 Tautenburg, Germany.

light sources. The LIGHTMETER is fully weatherproof, maintenance-free, operates during the whole year, and is sensitive down to the darkest night time conditions. Thus, it is the perfect device to establish a large network all over the planet in order to detect changes and measure long-term trends in the night sky brightness caused by growing cities and industry (light pollution). In addition, the data of LIGHT-METER can support site selection processes and operation of astronomical observing facilities (§ 3).

## 2. OBSERVATION AND DATA REDUCTION

We present a first set of measurements taken at Cerro Armazones in Chile, the site of the future E-ELT. The data set covers 107 days, starting from January 28th, 2010 to May 14th, 2010. The LIGHTMETER was mounted next to the platform of the DIMM tower (differential image motion monitor, Sarazin & Roddier 1990) and was operated with a data sampling rate of 1 Hz. A text file is produced as output, containing date, time, measured raw counts, and temperature from an internal temperature sensor.

In order to be able to measure over the whole natural brightness range, the LIGHTMETER has a non-linear behavior. By construction non-linearity typically starts at full moon brightness, with a percent level departure from a linear relation, say. That maximizes the linear range available for dark-sky measurements. In addition, the sensitivity of the sensor can degrade with time and its sensitivity is temperature dependent. To account for these variable effects we are using the Sun and the Moon as natural calibrator sources.

To reduce the huge amount of data (86 400 measurements per day), we are using a pipeline written in the Interactive Data Language (IDL). The pipeline performs several steps which are summarized in the following:

- sorting the data for each day (one day is equal to one Julian day)
- detection of clouds by computing a running root mean square (RMS)
- computation of Lunar and Solar ephemerids
- computation of the theoretical Lunar irradiance by using a disk-equivalent reflectance model by Kieffer & Stone (2005) and ASTM E490 Air Mass Zero solar spectral irradiance data (ASTM E490)<sup>4</sup>

- computation of the theoretical Solar irradiance by using ASTM E490
- for the computation of the airmass the model by Rozenberg (1966)
- for the computation of the extinction caused by atmosphere of the Earth the model by Hayes & Latham (1975) is used
- the spectral response of the LIGHTMETER is taken into account
- simultaneous fit of the LIGHTMETER raw counts to the computed irradiance model (Sun and Moon), only parts without clouds and altitudes  $\geq +20^{\circ}$  of Sun and Moon are considered

The fit of the data to the irradiance model is done using following expression:

$$I = c \cdot \left\{ b \cdot \left[ a \cdot \exp\left(\frac{cnt \cdot (1 + d \cdot T)}{a}\right) - 1 \right] \right\}$$
  
+  $c \cdot cnt \cdot (1 + d \cdot T),$  (1)

where I is the measured irradiance by the Light-Meter, a, b, c, and d are the fit parameters, T is the measured temperature of the Lightmeter, and cnt are the measured raw counts. This expression accounts for the non-linearity behavior as well as the temperature dependent sensitivity of the Lightmeter. This fit is carried out on a daily basis using following constraints:

- the altitude of Sun and Moon have to be  $\geq +20^{\circ}$
- sky conditions have to be clear

Therefore, we directly account for degradation effects in sensitivity of the device, too. The upper plot of Figure 2 shows an example of the fit of Light-METER measurements to the computed irradiance model for March 29th, 2010 (no clouds were present, night with full moon). Measurements that are not considered for the fit are not plotted (e.g. twilight, low altitudes of Sun and Moon). The lower plot of Figure 2 shows the relative RMS of the fit between the irradiance model and the measured data. The scatter in the relative RMS around noon (15 to 18 p.m. UT) is caused by the heating of the Light-METER due to solar radiation. The temperature of the sensor easily reaches 50°C during noon at Cerro Armazones and Equation 1 does not account for additional side-effects of the sensor that might occur during this time. The pipeline produces several files as output. A daily night log is produced containing

 $<sup>^4 \</sup>rm http://rredc.nrel.gov/solar/spectra/am0$ 




Fig. 2. Example of a fit of the measured LIGHT-METER data to the irradiance model for March 29th, 2010 (no clouds were present, night with full moon). The relative RMS (lower plot) of the fit between the measured data and the irradiance model inside the range of 10%. Because of the good fit performance the measured data are not distinguishable from the computed irradiance model.

the main measurements of the night: date, length of the night, illuminated fraction of the moon, JD, time stamps in UT, Irradiance, data of the running RMS (cloud detection), temperature of the sensor, and altitude of the moon. The fit parameters are written into a file and get updated after each successful fit. The measured irradiance of the day and night is written into a file and plotted. Additional plots and files are generated like the median night sky irradiance and the proportion of clear time during the night for statistical purpose.

# 3. FIRST RESULTS

The continuous measurement of the night sky brightness gives us a detailed view on the conditions of the observing site. Short-term variations caused by, e.g. clouds, and long-term trends caused by, e.g. the seasonal appearance of the Milky Way, can be identified. Figure 3 shows the average night sky irradiance of Cerro Armazones derived from our data set. Only dark time was considered for this computation, i.e., only parts where the altitude of the Sun was  $\leq -18^{\circ}$  and where the altitude of the Moon was  $\leq -5^{\circ}$  were considered. In addition, the first two hours after the end and the last two hours before the beginning of the astronomical twilight were rejected from the computation to avoid contamination by the Zodiacal light. Only nights that fulfill




Fig. 3. Median night sky irradiance for Cerro Armazones.

all constraints are plotted. The error bars are the standard deviation of the measured irradiance during the night and are an indicator for the stability of the night sky irradiance, i.e., as smaller the error bar as stable the night sky irradiance. The change of the irradiance over time is caused by the seasonal appearance and disappearance of the Milky Way.

Figure 4 shows a histogram of the measured irradiance during the night of our data set obtained at Cerro Armazones. Only data are considered where the altitude of the Sun was  $\leq -18^{\circ}$ . There are three distinct features visible: 1.) The highest peak on the left side represents dark and clear nights. As narrower the peak as better the overall sky quality of the observing site, i.e. it is an indicator for the presence of clouds and artificial light sources. 2.) Next to the first peak a smaller peak is present at a higher irradiance (log  $I = -5.1 \text{ W} \text{ m}^{-2}$ ). This is the contribution of the Milky Way, Zodiacal light, and sky glow. 3.) The broad peak at the very right, around  $\log I = -3.5 \text{ W m}^{-2}$ , is the contribution by the Moon. These features are not visible in light polluted areas. In such a case, the histogram would only show a single broad peak and clear nights would no longer distinguishable from nights where full moon is present.

# 4. CONCLUSION

The LIGHTMETER is a new sensitive device, which allows continuous measurements of the night sky brightness independent from the observing site and weather conditions. It covers the whole brightness range from day time down to the darkest night time conditions. The usage of the Sun and the Moon as natural calibrators keeps the hardware requirements low. We presented a first data set taken at Cerro Armazones and proofed the high sensitivity of the LIGHTMETER. The high data sampling rate of




Fig. 4. Histogram of the measured irradiance during the night.

 $1~\mathrm{Hz}$  allows real-time measurements of the night sky brightness

which can be used to support astronomical observations, e.g. for cloud detection, and can contribute to site selection processes of future telescope facilities.

# REFERENCES

Hayes, D. S., & Latham, D. W. 1975, ApJ, 197, 593
Kieffer, H. H., & Stone, T. C. 2005, AJ, 129, 2887
Rozenberg, G. V. 1966, Twilight: A Study in Atmospheric Optics. New York: Plenum Press, 160
Sarazin, M., & Roddier, F. 1990, A&A, 227, 294